



LAWRENCE
LIVERMORE
NATIONAL
LABORATORY

Chemical imaging of biological materials by NanoSIMS using isotopic and elemental labels

P. K. Weber, S. J. Fallon, J. Pett-Ridge, S.
Ghosal, I. D. Hutcheon

April 10, 2006

19th Annual Workshop on SIMS
Rancho Mirage, CA, United States
May 16, 2006 through May 19, 2006

This document was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor the University of California nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or the University of California, and shall not be used for advertising or product endorsement purposes.

Chemical imaging of biological materials by NanoSIMS using isotopic and elemental labels

Peter K. Weber^{a, *}, Stewart J. Fallon^a, Jennifer Pett-Ridge^a, Sutapa Ghosal^a, and Ian D. Hutcheon^a

^aChemical Biology and Nuclear Science, 7000 East Ave., L-231, Lawrence Livermore National Laboratory,
Livermore, CA 94550, USA
Email: weber21@llnl.gov

The NanoSIMS 50 combines unprecedented spatial resolution (as good as 50 nm) with ultra-high sensitivity (minimum detection limit of ~200 atoms). The NanoSIMS 50 incorporates an array of detectors, enabling simultaneous collection of 5 species originating from the same sputtered volume of a sample. The primary ion beam (Cs^+ or O^-) can be scanned across the sample to produce quantitative secondary ion images. This capability for multiple isotope imaging with high spatial resolution provides a novel new approach to the study of biological materials. Studies can be made of sub-regions of tissues, mammalian cells, and bacteria. Major, minor and trace element distributions can be mapped on a submicron scale, growth and metabolism can be tracked using stable isotope labels, and biogenic origin can be determined based on composition. We have applied this technique extensively to mammalian and prokaryotic cells and bacterial spores. The NanoSIMS technology enables the researcher to interrogate the fate of molecules of interest within cells and organs through elemental and isotopic labeling. Biological applications at LLNL will be discussed.

This work was performed under the auspices of the U.S. Department of Energy by the University of California, Lawrence Livermore National Laboratory under Contract W-7405-Eng-48.